



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/719,305	11/21/2003	Robert H. Wham	2934	3007
50855	7590	03/02/2007	EXAMINER	
UNITED STATES SURGICAL, A DIVISION OF TYCO HEALTHCARE GROUP LP 195 MCDERMOTT ROAD NORTH HAVEN, CT 06473			PEFFLEY, MICHAEL F	
			ART UNIT	PAPER NUMBER
			3739	
SHORTENED STATUTORY PERIOD OF RESPONSE	MAIL DATE	DELIVERY MODE		
3 MONTHS	03/02/2007	PAPER		

Please find below and/or attached an Office communication concerning this application or proceeding.

If NO period for reply is specified above, the maximum statutory period will apply and will expire 6 MONTHS from the mailing date of this communication.

Office Action Summary	Application No.	Applicant(s)
	10/719,305	WHAM ET AL.
Examiner	Michael Peffley	Art Unit 3739

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

1) Responsive to communication(s) filed on 21 November 2003.

2a) This action is **FINAL**. 2b) This action is non-final.

3) Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

4) Claim(s) 1-7 is/are pending in the application.
4a) Of the above claim(s) _____ is/are withdrawn from consideration.

5) Claim(s) _____ is/are allowed.

6) Claim(s) 1-7 is/are rejected.

7) Claim(s) _____ is/are objected to.

8) Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

9) The specification is objected to by the Examiner.

10) The drawing(s) filed on 11-21-03 is/are: a) accepted or b) objected to by the Examiner.

Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).

Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).

11) The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

12) Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
a) All b) Some * c) None of:
1. Certified copies of the priority documents have been received.
2. Certified copies of the priority documents have been received in Application No. _____.
3. Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

1) Notice of References Cited (PTO-892)
2) Notice of Draftsperson's Patent Drawing Review (PTO-948)
3) Information Disclosure Statement(s) (PTO/SB/08)
Paper No(s)/Mail Date. *See Continuation Sheet.*

4) Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____

5) Notice of Informal Patent Application

6) Other: _____

Information Disclosure Statement

Applicant should note that the large number of references in the attached IDS have been considered by the examiner in the same manner as other documents in Office search files are considered by the examiner while conducting a search of the prior art in a proper field of search. **See MPEP 609.05(b).** Applicant is requested to point out any particular references in the IDS which they believe may be of particular relevance to the instant claimed invention in response to this office action.

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claims 1, 2 and 7 are rejected under 35 U.S.C. 102(b) as being anticipated by Klicek (6,210,403).

Klicek discloses a system for controlling an output of an electrosurgical generator comprising a drive circuit (33) responsive to a feedback signal, at least one sensing circuit (17,19) coupled to the patient electrode for generating at least one signal (18,20) corresponding to a value of a waveform present at the electrode, a processing circuit (21) for receiving the at least one signal, a determining circuit (24) in communication with the processing circuit for generating an output signal, and a control circuit (27) for generating a feedback signal representative of the difference between a value of the output (25) and a reference value (26).

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 3-6 are rejected under 35 U.S.C. 103(a) as being unpatentable over Klicek ('403) in view of the teachings of Klicek (5,372,596) and Ogden ("Goertzel Alternative to the Fourier Transform").

The Klicek ('403) system has been previously addressed. It is substantially identical to the circuitry disclosed in applicant's specification. However, Klicek ('403) fails to disclose the use of a Goertzel Algorithm for determining a phase difference between waveforms to control the feedback to the drive source.

Klicek ('596) disclose another electrosurgical generator control system very much similar to the Klicek ('403) system. In particular, Klicek ('596) also disclose sensing voltage and current and using those signals in a control feedback loop just as in the Klicek ('403) system. Klicek ('596) further specifically teach that it is advantageous to provide phase difference comparisons to compensate for the energy delivery to the operating site (see Abstract and Figure 1). Klicek ('596) fail to specifically disclose the use of a Goertzel Algorithm to determine the frequency components of the signal. It is submitted that typically known Fourier Transform methods are used to determine the phase characteristics of the signal as is generally known in the art.

Ogden teaches of the use of the Goertzel Algorithm as a substitute to DFT to determine frequency characteristics of signals. Specifically, Ogden teaches that the Goertzel Algorithm is a faster, easier substitute for the DFT.


To have provided the Kliceck ('403) system with a phase difference compensation circuitry to provide more accurate feedback to the electrosurgical generator would have been an obvious modification for one of ordinary skill in the art in view of the teaching of Kliceck ('596). To have further used the Goertzel Algorithm to provide frequency response characteristics to the measured signals would have been an obvious consideration for one of ordinary skill in the art since Ogden fairly teaches that this is a known substitute for DFT processing that is much quicker and simpler.

Conclusion

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Michael Peffley whose telephone number is (571) 272-4770. The examiner can normally be reached on Mon-Fri from 6am-3pm.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Linda Dvorak can be reached on (571) 272-4764. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.


Michael Peffley
Primary Examiner
Art Unit 3739

mp
February 28, 2007

Continuation of Attachment(s) 3). Information Disclosure Statement(s) (PTO/SB/08), Paper No(s)/Mail Date :3/24/04; 9/16/04; 10/24/05; 5/5/06. (1124105- Copy of 3194105)